
Midmark M11 - v1.0.5
Total Cycles: 350

Name: RITTER2
Sterilizer ID: V2313228
09/09/2021 03:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.2 F 0.0
2:00 188.2 F 4.1
4:00 199.9 F 9.4
6:00 224.3 F 7.3
8:00 230.0 F 11.3
10:00 236.3 F 9.6
12:00 237.9 F 9.3
14:00 249.9 F 15.7
16:00 262.0 F 22.7
18:00 272.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.6
1:00 271.9 F 29.8
2:00 271.9 F 29.8
3:00 271.8 F 29.7
4:00 271.9 F 29.5

Min 271.8 F 29.5
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 226.8 F 0.0
10:00 222.3 F 0.0
15:00 222.8 F 0.0
20:00 218.1 F 0.0
25:00 219.7 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 18:55
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:56:56
CYCLE COMPLETE

09/09/2021 04:14 PM
ACCEPTED: X
REJECTED:
Operator: U02