
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
1.8uS / 1.1ppm
CYCLE NUMBER 000041
17:06 10/08/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:21
PRESSURIZING 2:35
STERILIZING 3:28
136.0 C 222kPa 3:28
136.3 C 226kPa 3:58
136.3 C 226kPa 4:29
136.5 C 227kPa 4:59
136.4 C 227kPa 5:29
136.5 C 228kPa 5:59
136.4 C 224kPa 6:29
136.5 C 228kPa 6:58
VENTING 6:59
AIR DRYING 7:54
CYCLE COMPLETE 37:54

Digital Signature #
810105EFB80DA9B2
