
Midmark M11 - v1.0.5
Total Cycles: 2463

Name: RITTER2
Sterilizer ID: V2313228
06/05/2023 10:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.0 F 0.0
2:00 189.1 F 3.1
4:00 194.4 F 8.4
6:00 214.8 F 10.1
8:00 227.6 F 9.9
10:00 232.6 F 8.4
12:00 242.7 F 12.6
14:00 244.0 F 12.9
16:00 256.7 F 19.7
18:00 267.8 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.9 F 29.9
2:00 271.7 F 29.3
3:00 271.3 F 29.4
4:00 273.4 F 30.2

Min 271.0 F 29.0
Max 273.5 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 215.3 F 0.0
10:00 226.6 F 0.0
15:00 227.0 F 0.0
20:00 226.0 F 0.0
25:00 228.8 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 19:21
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

06/05/2023 11:36 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____