
Midmark M11 - v1.0.0
Total Cycles: 2255

Name: RITTER 1
Sterilizer ID: V2178198
08/29/2022 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.6 F 0.0
2:00 179.5 F 0.2
4:00 185.7 F 0.8
6:00 205.3 F 4.0
8:00 221.0 F 8.8
10:00 234.2 F 12.1
12:00 237.1 F 10.4
14:00 234.2 F 8.0
16:00 241.1 F 10.7
18:00 247.4 F 13.8
20:00 259.7 F 20.5
22:00 270.7 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.6 F 28.9
2:00 271.9 F 28.2
3:00 271.9 F 28.5
4:00 272.7 F 28.9

Min 271.5 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 229.5 F 0.0
10:00 226.5 F 0.0
15:00 225.6 F 0.0
20:00 229.5 F 0.0
25:00 229.2 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 22:51
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:02
CYCLE COMPLETE

08/29/2022 09:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____