

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2351

Name: RITTER2  
Sterilizer ID: V2313228  
05/10/2023 02:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.9 F 0.0  
2:00 184.5 F 3.6  
4:00 197.0 F 8.4  
6:00 221.4 F 10.2  
8:00 227.2 F 10.0  
10:00 232.5 F 8.2  
12:00 242.1 F 12.4  
14:00 244.2 F 12.9  
16:00 256.5 F 19.6  
18:00 267.7 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.3  
1:00 272.4 F 30.1  
2:00 272.0 F 29.4  
3:00 271.8 F 29.9  
4:00 272.3 F 29.7  
  
Min 271.2 F 29.1  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.7 F 0.0  
5:00 211.5 F 0.0  
10:00 227.2 F 0.0  
15:00 229.2 F 0.0  
20:00 229.1 F 0.0  
25:00 228.4 F 0.0  
30:00 227.0 F 0.0  
Drying Complete

FILLING: 2:21  
HEATING: 19:13  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:50  
CYCLE COMPLETE

05/10/2023 03:34 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_