
Midmark M11 - v1.0.5
Total Cycles: 402

Name: RITTER2
Sterilizer ID: V2313228
09/28/2021 01:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.1 F 0.0
2:00 154.9 F 2.8
4:00 170.8 F 7.5
6:00 199.5 F 13.2
8:00 214.6 F 8.3
10:00 230.8 F 11.1
12:00 236.1 F 12.2
14:00 241.3 F 11.2
16:00 251.7 F 18.0
18:00 264.9 F 25.0
20:00 272.9 F 30.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.8
1:00 271.7 F 29.9
2:00 271.8 F 29.9
3:00 271.6 F 29.8
4:00 271.8 F 29.7

Min 271.6 F 29.7
Max 274.7 F 31.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.2 F 0.0
5:00 228.8 F 0.0
10:00 225.2 F 0.0
15:00 233.8 F 0.0
Dry Cycle Aborted

FILLING: 1:39
HEATING: 20:16
STERILIZING: 4:00
VENTING: 3:04
DRYING: 19:41
TOTAL CYCLE: 00:48:40

09/28/2021 01:54 PM
ACCEPTED: X
REJECTED:
Operator: U01