
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002629
12:53 08/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:55
STERILIZING 3:46
136.0 C 221kPa 3:46
136.4 C 223kPa 4:16
136.4 C 222kPa 4:46
136.4 C 224kPa 5:16
136.4 C 224kPa 5:46
136.2 C 222kPa 6:16
136.4 C 224kPa 6:46
136.2 C 221kPa 7:16
136.2 C 221kPa 7:16
VENTING 7:16
AIR DRYING 8:10
CYCLE COMPLETE 8:11

Digital Signature #
55D51EF464F00938
