

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2141

Name: RITTER 1  
Sterilizer ID: V2178198  
07/18/2022 07:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.0 F 0.0  
2:00 72.7 F 0.0  
4:00 79.5 F 0.0  
6:00 105.9 F 0.0  
8:00 136.1 F 0.1  
10:00 170.6 F 2.0  
12:00 189.3 F 5.2  
14:00 202.1 F 8.3  
16:00 214.8 F 11.7  
18:00 221.6 F 5.1  
20:00 233.0 F 9.6  
22:00 238.9 F 10.9  
24:00 240.6 F 11.1  
26:00 236.4 F 8.9  
28:00 250.3 F 15.3  
30:00 261.2 F 21.4  
32:00 271.2 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.1  
1:00 272.5 F 28.9  
2:00 271.7 F 28.2  
3:00 271.5 F 28.5  
4:00 272.4 F 28.8  
  
Min 271.1 F 28.0  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.5 F 0.0  
5:00 231.4 F 0.0  
10:00 224.0 F 0.0  
15:00 228.2 F 0.0  
20:00 234.3 F 0.0  
25:00 230.9 F 0.0  
30:00 226.9 F 0.0  
  
Drying Complete

FILLING: 0:46  
HEATING: 32:47  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:10:42  
CYCLE COMPLETE

07/18/2022 08:48 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_