
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001401
10:21 05/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:11
PRESSURIZING 4:12
STERILIZING 5:28
136.0 C 220kPa 5:28
136.4 C 223kPa 5:58
136.2 C 223kPa 6:28
136.4 C 225kPa 6:58
136.5 C 223kPa 7:28
136.3 C 222kPa 7:58
136.4 C 224kPa 8:28
136.1 C 223kPa 8:58
136.1 C 223kPa 8:58
VENTING 8:58
AIR DRYING 9:55
CYCLE COMPLETE 24:55

Digital Signature #
C1617CDFF543762D
