

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3180

Name: RITTER 1  
Sterilizer ID: V2178198  
05/01/2023 09:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 209.0 F 0.0  
2:00 198.9 F 2.8  
4:00 209.9 F 7.5  
6:00 227.9 F 13.2  
8:00 230.1 F 8.6  
10:00 240.6 F 11.0  
12:00 241.9 F 12.1  
14:00 240.4 F 11.0  
16:00 254.2 F 17.9  
18:00 266.0 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.4 F 29.2  
2:00 271.7 F 29.1  
3:00 272.4 F 29.2  
4:00 271.4 F 28.8  
  
Min 271.4 F 28.6  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.2 F 0.0  
5:00 217.9 F 0.0  
10:00 236.2 F 0.0  
15:00 234.4 F 0.0  
20:00 228.6 F 0.0  
25:00 223.9 F 0.0  
30:00 224.5 F 0.0  
Drying Complete

FILLING: 1:46  
HEATING: 19:37  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:58:42  
CYCLE COMPLETE

05/01/2023 10:53 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_