
Midmark M11 - v1.0.5
Total Cycles: 3463

Name: RITTER2
Sterilizer ID: V2313228
03/21/2024 08:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 70.3 F 0.0
4:00 77.9 F 0.0
6:00 112.6 F 1.1
8:00 150.2 F 3.9
10:00 168.2 F 7.4
12:00 191.4 F 11.4
14:00 209.2 F 4.1
16:00 223.5 F 7.6
18:00 233.9 F 11.5
20:00 230.0 F 7.3
22:00 241.6 F 12.6
24:00 239.5 F 10.6
26:00 252.2 F 16.8
28:00 262.8 F 23.1
30:00 272.7 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.7 F 29.8
2:00 272.1 F 29.0
3:00 271.7 F 29.3
4:00 272.7 F 29.6

Min 271.1 F 28.7
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 229.5 F 0.0
10:00 229.3 F 0.0
15:00 177.3 F 0.0
20:00 172.2 F 0.0
25:00 169.8 F 0.0
30:00 168.6 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:25
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:08:29
CYCLE COMPLETE

03/21/2024 09:44 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____