

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4852

Name: RITTER 1  
Sterilizer ID: V2178198  
08/26/2024 02:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 81.9 F 0.0  
2:00 84.2 F 0.0  
4:00 116.1 F 0.3  
6:00 147.3 F 1.2  
8:00 168.7 F 2.1  
10:00 186.4 F 4.7  
12:00 198.9 F 7.8  
14:00 207.9 F 11.2  
16:00 221.6 F 5.2  
18:00 226.3 F 8.0  
20:00 236.7 F 12.2  
22:00 235.3 F 8.9  
24:00 242.7 F 11.5  
26:00 244.7 F 12.5  
28:00 256.0 F 18.3  
30:00 265.8 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.1 F 28.7  
2:00 272.3 F 28.7  
3:00 271.8 F 28.2  
4:00 271.3 F 28.1  
  
Min 271.2 F 28.0  
Max 272.4 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.0 F 0.0  
5:00 194.1 F 0.0  
10:00 198.3 F 0.0  
15:00 199.7 F 0.0  
20:00 216.8 F 0.0  
25:00 232.7 F 0.0  
30:00 231.1 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:56  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:10:32  
CYCLE COMPLETE

08/26/2024 03:40 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_