

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3069

Name: RITTER 1  
Sterilizer ID: V2178198  
04/03/2023 02:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.2 F 0.0  
2:00 200.1 F 3.6  
4:00 218.6 F 8.9  
6:00 226.7 F 7.6  
8:00 235.6 F 10.8  
10:00 236.0 F 8.5  
12:00 240.1 F 9.8  
14:00 248.6 F 14.7  
16:00 261.6 F 22.0  
18:00 272.8 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.2 F 28.8  
2:00 272.7 F 29.3  
3:00 271.6 F 28.5  
4:00 272.5 F 29.1  
  
Min 271.6 F 28.4  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.3 F 0.0  
5:00 233.3 F 0.0  
10:00 229.3 F 0.0  
15:00 226.8 F 0.0  
20:00 229.3 F 0.0  
25:00 230.1 F 0.0  
30:00 227.9 F 0.0  
  
Drying Complete

FILLING: 2:10  
HEATING: 18:17  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:57:41  
CYCLE COMPLETE

04/03/2023 03:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_