

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.7uS / 0.4ppm  
CYCLE NUMBER 001856  
15:18 15/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:21  
PRESSURIZING 2:26  
STERILIZING 3:31  
136.4 C 225kPa 3:31  
136.4 C 226kPa 4:01  
136.4 C 225kPa 4:31  
136.4 C 226kPa 5:01  
136.4 C 224kPa 5:31  
136.1 C 223kPa 6:01  
136.4 C 226kPa 6:31  
136.3 C 226kPa 7:01  
136.3 C 226kPa 7:01  
VENTING 7:01  
AIR DRYING 8:00  
CYCLE COMPLETE 8:25

-----  
Digital Signature #  
CC846413E0A57E21  
-----