
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001317
12:33 02/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:04
STERILIZING 4:05
136.4 C 223kPa 4:05
136.1 C 222kPa 4:35
136.5 C 225kPa 5:05
136.2 C 220kPa 5:35
136.3 C 223kPa 6:05
136.4 C 224kPa 6:35
136.0 C 221kPa 7:05
136.4 C 222kPa 7:35
136.4 C 222kPa 7:35
VENTING 7:36
AIR DRYING 8:33
CYCLE COMPLETE 8:39

Digital Signature #
25B944D10CAA8F9E
