
Midmark M11 - v1.0.5
Total Cycles: 1681

Name: RITTER2
Sterilizer ID: V2313228
11/03/2022 02:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.1 F 0.0
2:00 160.6 F 2.7
4:00 164.8 F 6.4
6:00 188.1 F 11.3
8:00 216.4 F 5.8
10:00 227.6 F 11.3
12:00 231.5 F 8.2
14:00 240.7 F 11.6
16:00 244.1 F 12.8
18:00 256.3 F 19.3
20:00 267.4 F 26.2
22:00 272.9 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.6
1:00 272.6 F 30.3
2:00 272.5 F 29.8
3:00 272.0 F 29.8
4:00 272.8 F 30.1

Min 271.7 F 29.4
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 219.3 F 0.0
10:00 238.2 F 0.0
15:00 227.8 F 0.0
20:00 224.5 F 0.0
25:00 223.7 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 22:06
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:39
CYCLE COMPLETE

11/03/2022 03:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____