

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000690  
11:07 15/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:25  
PRESSURIZING 2:30  
STERILIZING 3:33  
136.0 C 228kPa 3:33  
136.1 C 232kPa 4:03  
136.3 C 232kPa 4:33  
135.9 C 229kPa 5:03  
136.4 C 232kPa 5:33  
136.2 C 231kPa 6:03  
136.3 C 233kPa 6:33  
136.3 C 232kPa 7:03  
136.3 C 232kPa 7:03  
VENTING 7:03  
AIR DRYING 7:55  
CYCLE COMPLETE 7:58  
-----

Digital Signature #  
C7E27208B24E885E  
-----