
Midmark M11 - v1.0.5
Total Cycles: 3878

Name: RITTER2
Sterilizer ID: V2313228
08/21/2024 11:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.6 F 0.0
2:00 160.7 F 2.5
4:00 169.2 F 5.7
6:00 196.6 F 10.7
8:00 215.5 F 4.9
10:00 227.6 F 10.2
12:00 232.1 F 8.0
14:00 243.5 F 13.2
16:00 243.8 F 12.7
18:00 256.7 F 19.5
20:00 268.1 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.6 F 29.7
2:00 271.5 F 29.0
3:00 272.8 F 29.9
4:00 272.0 F 29.1

Min 271.5 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 217.8 F 0.0
10:00 228.0 F 0.0
15:00 228.8 F 0.0
20:00 225.7 F 0.0
25:00 224.5 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 21:16
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

08/21/2024 12:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____