

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3482

Name: RITTER2  
Sterilizer ID: V2313228  
03/27/2024 12:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.6 F 0.0  
2:00 131.6 F 0.8  
4:00 152.3 F 3.3  
6:00 177.4 F 6.9  
8:00 196.4 F 10.9  
10:00 213.7 F 5.1  
12:00 219.7 F 8.8  
14:00 232.9 F 13.4  
16:00 236.5 F 10.5  
18:00 235.3 F 8.7  
20:00 248.7 F 15.3  
22:00 260.0 F 21.7  
24:00 270.5 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.3  
1:00 272.6 F 30.1  
2:00 271.8 F 29.5  
3:00 271.1 F 29.4  
4:00 272.2 F 30.2

Min 271.0 F 29.2  
Max 273.0 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.1 F 0.0  
5:00 223.7 F 0.0  
10:00 223.6 F 0.0  
15:00 219.2 F 0.0  
20:00 219.7 F 0.0  
25:00 234.2 F 0.0  
30:00 239.6 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 24:53  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:03:40  
CYCLE COMPLETE

03/27/2024 01:56 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_