
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000496
14:21 21/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:34
STERILIZING 4:36
136.0 C 220kPa 4:36
136.1 C 223kPa 5:07
136.4 C 225kPa 5:37
136.3 C 222kPa 6:07
136.4 C 224kPa 6:37
136.3 C 222kPa 7:07
136.4 C 221kPa 7:37
136.4 C 225kPa 8:07
136.4 C 225kPa 8:07
VENTING 8:07
AIR DRYING 9:02
CYCLE COMPLETE 9:04

Digital Signature #
47D4426AE96E5B1C
