

-----  
Midmark M11 - v1.0.5  
Total Cycles: 343

Name: RITTER2  
Sterilizer ID: V2313228  
09/09/2021 07:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.0 F 0.0  
2:00 77.2 F 0.0  
4:00 78.1 F 0.1  
6:00 118.7 F 0.3  
8:00 151.3 F 2.4  
10:00 176.2 F 6.0  
12:00 196.1 F 9.8  
14:00 211.5 F 9.8  
16:00 220.5 F 6.9  
18:00 232.0 F 10.9  
20:00 232.2 F 7.7  
22:00 242.2 F 12.3  
24:00 239.5 F 10.6  
26:00 252.0 F 16.7  
28:00 262.4 F 22.8  
30:00 272.2 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.3 F 29.2  
2:00 272.4 F 29.7  
3:00 271.7 F 28.8  
4:00 272.4 F 29.6

Min 271.5 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.6 F 0.0  
5:00 207.5 F 0.0  
10:00 206.4 F 0.0  
15:00 209.5 F 0.0  
20:00 184.7 F 0.0  
25:00 178.4 F 0.0  
30:00 169.1 F 0.0  
Drying Complete

FILLING: 1:21  
HEATING: 30:26  
STERILIZING: 4:00  
VENTING: 2:60  
DRYING: 30:00  
TOTAL CYCLE: 01:08:47  
CYCLE COMPLETE

09/09/2021 08:54 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02