
Midmark M11 - v1.0.0
Total Cycles: 1859

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2022 07:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.2 F 0.0
2:00 71.8 F 0.0
4:00 85.2 F 0.0
6:00 115.9 F 0.8
8:00 145.8 F 2.4
10:00 168.7 F 4.5
12:00 191.1 F 7.3
14:00 205.0 F 11.0
16:00 223.8 F 4.9
18:00 227.3 F 7.7
20:00 238.0 F 11.8
22:00 234.4 F 8.2
24:00 245.9 F 12.6
26:00 243.7 F 11.9
28:00 255.7 F 18.2
30:00 266.4 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.2 F 28.4
2:00 272.6 F 28.9
3:00 271.4 F 28.3
4:00 271.8 F 28.4

Min 271.2 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.3 F 0.0
5:00 193.0 F 0.0
10:00 199.1 F 0.0
15:00 194.6 F 0.0
20:00 195.9 F 0.0
25:00 197.3 F 0.0
30:00 207.3 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 31:36
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:09:26
CYCLE COMPLETE

04/25/2022 08:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____