
Midmark M11 - v1.0.5
Total Cycles: 1304

Name: RITTER2
Sterilizer ID: V2313228
07/15/2022 09:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.4 F 0.0
2:00 186.0 F 0.3
4:00 193.8 F 3.6
6:00 213.9 F 8.5
8:00 226.4 F 11.4
10:00 232.0 F 9.8
12:00 236.1 F 9.2
14:00 244.4 F 13.5
16:00 244.1 F 12.6
18:00 256.2 F 19.1
20:00 267.0 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.7 F 29.6
2:00 271.6 F 28.9
3:00 272.7 F 29.8
4:00 272.1 F 29.1

Min 271.5 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 231.3 F 0.0
10:00 228.2 F 0.0
15:00 224.8 F 0.0
20:00 226.3 F 0.0
25:00 229.9 F 0.0
30:00 229.9 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 21:33
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:00:10
CYCLE COMPLETE

07/15/2022 10:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____