
Midmark M11 - v1.0.0
Total Cycles: 1432

Name: RITTER 1
Sterilizer ID: V2178198
12/08/2021 11:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.2 F 0.0
2:00 198.4 F 2.7
4:00 199.9 F 7.4
6:00 218.9 F 13.4
8:00 228.8 F 8.0
10:00 241.8 F 13.3
12:00 240.0 F 10.7
14:00 236.5 F 8.8
16:00 251.2 F 15.7
18:00 263.1 F 22.6
20:00 271.5 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 271.9 F 28.2
2:00 272.6 F 28.7
3:00 271.7 F 28.4
4:00 271.9 F 28.3

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 226.2 F 0.0
10:00 216.8 F 0.0
15:00 212.4 F 0.0
20:00 200.2 F 0.0
25:00 199.9 F 0.0
30:00 200.9 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 20:04
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:58:51
CYCLE COMPLETE

12/08/2021 12:26 PM
ACCEPTED: X
REJECTED:
Operator: U02