
Midmark M11 - v1.0.0
Total Cycles: 2691

Name: RITTER 1
Sterilizer ID: V2178198
12/19/2022 01:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 213.7 F 0.0
2:00 205.0 F 3.7
4:00 217.8 F 8.8
6:00 227.6 F 7.1
8:00 235.7 F 10.6
10:00 235.1 F 8.0
12:00 241.3 F 10.3
14:00 247.5 F 14.0
16:00 260.3 F 21.3
18:00 272.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.5 F 29.0
2:00 271.5 F 28.6
3:00 272.6 F 29.2
4:00 271.8 F 28.5

Min 271.2 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.4 F 0.0
5:00 242.3 F 0.0
10:00 232.4 F 0.0
15:00 229.0 F 0.0
20:00 231.3 F 0.0
25:00 229.3 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 18:30
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:45
CYCLE COMPLETE

12/19/2022 02:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____