

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001156  
9:13 08/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:51  
PRESSURIZING 3:53  
STERILIZING 5:02  
136.0 C 220kPa 5:02  
136.2 C 223kPa 5:32  
136.4 C 224kPa 6:02  
136.4 C 224kPa 6:32  
136.4 C 224kPa 7:02  
136.3 C 224kPa 7:32  
136.4 C 225kPa 8:02  
136.2 C 223kPa 8:32  
136.2 C 223kPa 8:32  
VENTING 8:32  
AIR DRYING 9:26  
CYCLE COMPLETE 24:26  
-----

Digital Signature #  
321486E13F5F6E1D  
-----