
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000378
14:09 21/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:42
STERILIZING 4:11
136.5 C 233kPa 4:11
136.3 C 233kPa 4:41
136.4 C 231kPa 5:11
136.1 C 231kPa 5:41
136.3 C 231kPa 6:11
136.4 C 233kPa 6:41
136.4 C 232kPa 7:11
136.1 C 231kPa 7:41
136.1 C 231kPa 7:41
VENTING 7:41
AIR DRYING 8:36
CYCLE COMPLETE 13:01

Digital Signature #
1B73F29ED8B2D523
