
Midmark M11 - v1.0.5
Total Cycles: 666

Name: RITTER2
Sterilizer ID: V2313228
12/29/2021 02:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.8 F 0.0
2:00 175.1 F 3.9
4:00 190.0 F 8.7
6:00 213.1 F 8.7
8:00 226.9 F 9.5
10:00 234.0 F 9.2
12:00 242.8 F 13.4
14:00 244.0 F 12.7
16:00 257.0 F 19.5
18:00 268.2 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.0
1:00 272.2 F 29.4
2:00 271.6 F 29.4
3:00 272.0 F 29.3
4:00 271.3 F 29.3

Min 271.0 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.9 F 0.0
5:00 216.8 F 0.0
10:00 236.6 F 0.0
15:00 226.4 F 0.0
20:00 221.9 F 0.0
25:00 221.8 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:15
CYCLE COMPLETE

12/29/2021 03:44 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____