
Midmark M11 - v1.0.0
Total Cycles: 5275

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2025 02:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 97.8 F 0.0
2:00 98.5 F 0.9
4:00 123.3 F 3.2
6:00 144.9 F 5.9
8:00 208.8 F 8.8
10:00 203.3 F 12.9
12:00 213.8 F 3.7
14:00 220.2 F 7.1
16:00 231.8 F 10.9
18:00 232.5 F 7.0
20:00 241.2 F 11.0
22:00 235.4 F 7.7
24:00 248.1 F 13.9
26:00 258.5 F 19.6
28:00 268.3 F 25.8
30:00 272.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.6 F 29.0
2:00 271.8 F 28.3
3:00 272.6 F 29.1
4:00 272.3 F 28.6

Min 271.5 F 28.3
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.3 F 0.0
5:00 237.0 F 0.0
10:00 225.5 F 0.0
15:00 192.8 F 0.0
20:00 221.2 F 0.0
25:00 190.9 F 0.0
30:00 183.0 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 30:10
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:08:50
CYCLE COMPLETE

01/24/2025 03:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____