

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 007740  
13:57 13/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:09  
PRESSURIZING 3:10  
STERILIZING 4:13  
136.4 C 222kPa 4:13  
136.4 C 225kPa 4:43  
136.4 C 224kPa 5:13  
135.9 C 220kPa 5:43  
136.2 C 223kPa 6:13  
136.4 C 223kPa 6:43  
136.4 C 226kPa 7:13  
136.2 C 222kPa 7:43  
136.2 C 222kPa 7:43  
VENTING 7:43  
AIR DRYING 8:40  
CYCLE COMPLETE 9:50

-----  
Digital Signature #  
F3C4755B9F880DCC  
-----