
Midmark M11 - v1.0.0
Total Cycles: 4862

Name: RITTER 1
Sterilizer ID: V2178198
08/29/2024 09:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.3 F 0.0
2:00 169.4 F 0.2
4:00 171.1 F 0.6
6:00 194.5 F 2.5
8:00 206.1 F 6.5
10:00 219.6 F 10.7
12:00 226.6 F 6.5
14:00 235.0 F 10.6
16:00 234.8 F 7.7
18:00 246.2 F 13.5
20:00 244.3 F 12.1
22:00 256.3 F 18.4
24:00 266.7 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 271.6 F 28.3
2:00 272.2 F 28.6
3:00 271.6 F 28.1
4:00 271.2 F 28.1

Min 271.0 F 27.8
Max 272.3 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 233.8 F 0.0
10:00 222.3 F 0.0
15:00 215.8 F 0.0
20:00 236.6 F 0.0
25:00 238.3 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 25:35
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:04:02
CYCLE COMPLETE

08/29/2024 10:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____