
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000044
9:35 27/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:33
STERILIZING 3:48
136.4 C 220kPa 3:48
136.1 C 222kPa 4:18
136.2 C 223kPa 4:48
136.1 C 222kPa 5:18
136.1 C 222kPa 5:48
136.2 C 223kPa 6:18
136.4 C 224kPa 6:48
136.4 C 225kPa 7:18
136.4 C 225kPa 7:18
VENTING 7:18
AIR DRYING 8:09
CYCLE COMPLETE 9:55

Digital Signature #
E2EC6C717536665E
