
Midmark M11 - v1.0.5
Total Cycles: 1864

Name: RITTER2
Sterilizer ID: V2313228
12/20/2022 02:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.4 F 0.0
2:00 195.1 F 4.4
4:00 200.6 F 9.8
6:00 219.3 F 6.2
8:00 231.2 F 11.7
10:00 236.1 F 9.6
12:00 237.2 F 8.9
14:00 251.1 F 16.0
16:00 263.3 F 23.4
18:00 271.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.4
1:00 272.8 F 29.9
2:00 271.9 F 29.4
3:00 272.9 F 30.0
4:00 272.1 F 29.4

Min 271.9 F 29.2
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 222.5 F 0.0
10:00 241.2 F 0.0
15:00 229.2 F 0.0
20:00 229.0 F 0.0
25:00 229.4 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 18:53
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:03
CYCLE COMPLETE

12/20/2022 02:59 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____