

-----  
Midmark M11 - v1.0.5  
Total Cycles: 185

Name: RITTER2  
Sterilizer ID: V2313228  
07/08/2021 02:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.3 F 0.0  
2:00 172.6 F 3.3  
4:00 196.5 F 8.3  
6:00 212.7 F 9.3  
8:00 229.5 F 9.5  
10:00 235.7 F 9.1  
12:00 243.4 F 12.1  
14:00 246.6 F 13.9  
16:00 259.4 F 20.9  
18:00 270.7 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.9  
1:00 271.9 F 28.8  
2:00 272.2 F 29.0  
3:00 272.5 F 29.4  
4:00 272.5 F 29.7  
  
Min 271.7 F 28.8  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.2 F 0.0  
5:00 227.3 F 0.0  
10:00 209.2 F 0.0  
15:00 211.2 F 0.0  
20:00 209.6 F 0.0  
25:00 205.9 F 0.0  
30:00 190.8 F 0.0  
Drying Complete

FILLING: 2:48  
HEATING: 18:39  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:36  
CYCLE COMPLETE

07/08/2021 03:21 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02