
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001500
9:34 05/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:16
STERILIZING 3:15
136.0 C 220kPa 3:15
136.1 C 223kPa 3:45
136.4 C 225kPa 4:15
136.1 C 222kPa 4:45
136.4 C 225kPa 5:15
136.1 C 223kPa 5:45
136.4 C 225kPa 6:15
136.1 C 223kPa 6:45
136.1 C 223kPa 6:45
VENTING 6:45
AIR DRYING 7:35
CYCLE COMPLETE 10:40

Digital Signature #
5CC989B60659C83E
