

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2479

Name: RITTER2  
Sterilizer ID: V2313228  
06/07/2023 11:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 202.9 F 0.0  
2:00 191.4 F 4.3  
4:00 199.8 F 9.3  
6:00 222.1 F 7.5  
8:00 228.0 F 10.2  
10:00 233.1 F 8.2  
12:00 242.8 F 12.1  
14:00 244.4 F 13.1  
16:00 257.0 F 19.9  
18:00 268.4 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.6 F 30.0  
2:00 271.7 F 29.2  
3:00 272.0 F 30.0  
4:00 272.4 F 29.8  
  
Min 271.3 F 29.1  
Max 272.6 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.4 F 0.0  
5:00 229.3 F 0.0  
10:00 229.4 F 0.0  
15:00 230.5 F 0.0  
20:00 230.3 F 0.0  
25:00 229.3 F 0.0  
30:00 227.3 F 0.0  
Drying Complete

FILLING: 1:56  
HEATING: 19:15  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:23  
CYCLE COMPLETE

06/07/2023 12:22 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_