
Midmark M11 - v1.0.0
Total Cycles: 4216

Name: RITTER 1
Sterilizer ID: V2178198
01/19/2024 12:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.7 F 0.0
2:00 144.9 F 2.5
4:00 166.2 F 6.0
6:00 188.5 F 10.3
8:00 220.5 F 5.2
10:00 223.8 F 8.9
12:00 235.6 F 10.9
14:00 238.8 F 10.7
16:00 235.4 F 8.5
18:00 250.0 F 15.1
20:00 261.3 F 21.6
22:00 271.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 273.0 F 29.1
2:00 271.3 F 28.3
3:00 272.5 F 29.2
4:00 272.3 F 28.8

Min 271.1 F 28.3
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 167.2 F 0.0
5:00 234.1 F 0.0
10:00 210.5 F 0.0
15:00 197.3 F 0.0
20:00 187.0 F 0.0
25:00 234.6 F 0.0
30:00 203.5 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:37
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:01:16
CYCLE COMPLETE

01/19/2024 01:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____