

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002791  
17:06 06/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:11  
PRESSURIZING 3:12  
STERILIZING 4:19  
136.4 C 224kPa 4:19  
136.4 C 226kPa 4:49  
136.4 C 220kPa 5:19  
136.4 C 223kPa 5:49  
136.4 C 224kPa 6:19  
136.4 C 224kPa 6:49  
136.4 C 224kPa 7:19  
136.4 C 223kPa 7:49  
136.4 C 223kPa 7:49  
VENTING 7:49  
AIR DRYING 8:45  
CYCLE COMPLETE 8:47

-----  
Digital Signature #  
CD74299AD3C57A44  
-----