

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4164

Name: RITTER2  
Sterilizer ID: V2313228  
11/28/2024 10:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.8 F 0.0  
2:00 71.6 F 0.0  
4:00 78.7 F 0.0  
6:00 112.9 F 0.1  
8:00 158.0 F 0.2  
10:00 183.1 F 0.5  
12:00 206.3 F 1.3  
14:00 218.7 F 4.6  
16:00 226.4 F 7.9  
18:00 233.5 F 11.6  
20:00 232.8 F 8.0  
22:00 243.9 F 13.2  
24:00 242.3 F 11.9  
26:00 240.9 F 11.0  
28:00 253.4 F 17.3  
30:00 264.2 F 23.8  
32:00 271.2 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.7  
1:00 272.7 F 29.3  
2:00 271.3 F 28.6  
3:00 272.6 F 29.4  
4:00 271.7 F 28.7  
  
Min 271.0 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.5 F 0.0  
5:00 233.1 F 0.0  
10:00 220.3 F 0.0  
15:00 232.9 F 0.0  
20:00 220.9 F 0.0  
25:00 226.1 F 0.0  
30:00 224.5 F 0.0  
  
Drying Complete

FILLING: 0:52  
HEATING: 32:04  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:10:36  
CYCLE COMPLETE

11/28/2024 11:50 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_