

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3378

Name: RITTER 1  
Sterilizer ID: V2178198  
06/13/2023 08:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 210.9 F 0.0  
2:00 198.3 F 0.0  
4:00 207.2 F 4.4  
6:00 223.9 F 9.8  
8:00 229.1 F 7.1  
10:00 237.4 F 11.1  
12:00 235.8 F 8.7  
14:00 239.7 F 9.9  
16:00 248.1 F 14.3  
18:00 260.6 F 21.2  
20:00 271.7 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.3  
1:00 272.6 F 29.1  
2:00 271.8 F 28.3  
3:00 272.7 F 29.1  
4:00 272.0 F 28.5  
  
Min 271.5 F 28.2  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.7 F 0.0  
5:00 223.3 F 0.0  
10:00 239.7 F 0.0  
15:00 227.4 F 0.0  
20:00 225.3 F 0.0  
25:00 229.9 F 0.0  
30:00 228.2 F 0.0  
Drying Complete

FILLING: 1:54  
HEATING: 20:27  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:59:36  
CYCLE COMPLETE

06/13/2023 09:27 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_