
Midmark M11 - v1.0.5
Total Cycles: 2607

Name: RITTER2
Sterilizer ID: V2313228
07/05/2023 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.3 F 0.0
2:00 139.7 F 0.0
4:00 147.8 F 1.1
6:00 185.3 F 5.1
8:00 206.1 F 9.8
10:00 219.0 F 7.8
12:00 226.9 F 9.5
14:00 234.9 F 9.8
16:00 240.4 F 11.9
18:00 238.7 F 10.3
20:00 251.9 F 16.7
22:00 263.1 F 23.3
24:00 272.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.9
1:00 271.7 F 29.4
2:00 272.5 F 29.4
3:00 271.3 F 28.8
4:00 272.1 F 29.7

Min 271.1 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 208.3 F 0.0
10:00 219.1 F 0.0
15:00 228.0 F 0.0
20:00 228.3 F 0.0
25:00 230.9 F 0.0
30:00 231.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 24:06
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:02:08
CYCLE COMPLETE

07/05/2023 09:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____