
Midmark M11 - v1.0.0
Total Cycles: 3105

Name: RITTER 1
Sterilizer ID: V2178198
04/13/2023 09:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.1 F 0.0
2:00 166.6 F 1.3
4:00 182.7 F 5.1
6:00 204.1 F 10.2
8:00 219.2 F 6.0
10:00 230.8 F 10.2
12:00 232.1 F 7.8
14:00 244.0 F 13.3
16:00 242.7 F 11.9
18:00 255.6 F 18.6
20:00 267.0 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.1
2:00 271.5 F 28.7
3:00 272.5 F 29.1
4:00 271.5 F 28.5

Min 271.4 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 213.5 F 0.0
10:00 228.9 F 0.0
15:00 218.4 F 0.0
20:00 215.7 F 0.0
25:00 216.7 F 0.0
30:00 215.0 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:27
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:57
CYCLE COMPLETE

04/13/2023 10:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____