
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000971
11:25 09/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:12
PRESSURIZING 2:17
STERILIZING 3:21
136.4 C 220kPa 3:21
136.3 C 233kPa 3:51
136.4 C 231kPa 4:21
136.3 C 233kPa 4:51
136.0 C 230kPa 5:21
136.2 C 232kPa 5:51
136.3 C 231kPa 6:21
136.3 C 233kPa 6:51
136.3 C 233kPa 6:51
VENTING 6:51
AIR DRYING 7:43
CYCLE COMPLETE 7:48

Digital Signature #
6231C3151BF71B28
