
Midmark M11 - v1.0.5
Total Cycles: 4011

Name: RITTER2
Sterilizer ID: V2313228
10/08/2024 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 84.6 F 0.0
2:00 84.7 F 0.0
4:00 89.6 F 0.0
6:00 120.7 F 0.2
8:00 155.7 F 0.4
10:00 182.1 F 2.2
12:00 194.1 F 5.3
14:00 206.0 F 8.5
16:00 218.6 F 12.1
18:00 224.7 F 7.4
20:00 235.3 F 12.1
22:00 235.7 F 9.6
24:00 239.4 F 10.1
26:00 246.5 F 13.9
28:00 257.2 F 19.7
30:00 266.9 F 25.9
32:00 273.1 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.5
1:00 272.4 F 29.5
2:00 272.7 F 29.7
3:00 272.6 F 30.0
4:00 271.9 F 29.8

Min 271.8 F 29.4
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.4 F 0.0
5:00 219.7 F 0.0
10:00 210.8 F 0.0
15:00 191.4 F 0.0
20:00 178.1 F 0.0
25:00 182.5 F 0.0
30:00 183.5 F 0.0

Drying Complete

FILLING: 0:53
HEATING: 32:20
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:11:02
CYCLE COMPLETE

10/08/2024 04:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____