

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000443  
7:05 15/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:37  
PRESSURIZING 4:38  
STERILIZING 6:01  
136.4 C 221kPa 6:01  
136.4 C 224kPa 6:31  
136.4 C 221kPa 7:01  
136.3 C 223kPa 7:31  
136.4 C 224kPa 8:01  
136.2 C 221kPa 8:31  
136.4 C 221kPa 9:01  
136.3 C 223kPa 9:31  
136.3 C 223kPa 9:31  
VENTING 9:31  
AIR DRYING 10:33  
CYCLE COMPLETE 10:41

-----  
Digital Signature #  
F5A4AF0F3BBB0378  
-----