

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4417

Name: RITTER2  
Sterilizer ID: V2313228  
03/10/2025 03:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 89.1 F 0.0  
2:00 89.0 F 0.4  
4:00 93.5 F 1.9  
6:00 120.8 F 4.5  
8:00 155.8 F 7.7  
10:00 184.6 F 11.0  
12:00 220.8 F 5.0  
14:00 207.7 F 7.5  
16:00 220.4 F 11.3  
18:00 229.3 F 6.5  
20:00 235.7 F 11.6  
22:00 236.0 F 9.2  
24:00 247.9 F 14.9  
26:00 258.7 F 20.7  
28:00 268.5 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 30.0  
1:00 271.8 F 30.2  
2:00 271.6 F 30.4  
3:00 271.6 F 30.4  
4:00 271.8 F 30.4  
  
Min 271.3 F 29.9  
Max 273.3 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.3 F 0.0  
5:00 231.5 F 0.0  
10:00 233.2 F 0.0  
15:00 219.8 F 0.0  
20:00 209.4 F 0.0  
25:00 202.5 F 0.0  
30:00 174.5 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 29:60  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:08:46  
CYCLE COMPLETE

03/10/2025 04:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_