
Midmark M11 - v1.0.5
Total Cycles: 1932

Name: RITTER2
Sterilizer ID: V2313228
01/12/2023 04:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.1 F 0.0
2:00 187.9 F 4.2
4:00 204.9 F 9.8
6:00 222.2 F 6.4
8:00 235.5 F 11.3
10:00 235.7 F 9.6
12:00 237.3 F 9.0
14:00 250.8 F 16.2
16:00 263.3 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.2
1:00 272.0 F 29.9
2:00 272.5 F 29.7
3:00 271.5 F 29.1
4:00 272.1 F 30.0

Min 271.3 F 29.0
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 246.5 F 0.0
10:00 229.2 F 0.0
15:00 226.9 F 0.0
20:00 229.0 F 0.0
25:00 224.3 F 0.0
30:00 225.4 F 0.0

Drying Complete

FILLING: 2:24
HEATING: 17:55
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:29
CYCLE COMPLETE

01/12/2023 05:52 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____