

-----  
Midmark M11 - v1.0.0  
Total Cycles: 691

Name: MIDMARK  
Sterilizer ID: V2178198  
02/23/2021 11:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.6 F 0.0  
2:00 140.6 F 1.1  
4:00 152.5 F 3.8  
6:00 182.1 F 7.9  
8:00 207.3 F 13.1  
10:00 226.7 F 6.8  
12:00 234.6 F 12.2  
14:00 237.3 F 9.3  
16:00 237.0 F 9.0  
18:00 247.8 F 14.1  
20:00 259.7 F 21.0  
22:00 270.7 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.3 F 29.0  
2:00 272.6 F 29.5  
3:00 271.8 F 28.7  
4:00 272.9 F 29.5

Min 271.2 F 28.6  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.6 F 0.0  
5:00 243.9 F 0.0  
10:00 238.9 F 0.0  
15:00 229.6 F 0.0  
20:00 215.6 F 0.0  
25:00 214.2 F 0.0  
30:00 211.5 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 22:42  
STERILIZING: 4:00  
VENTING: 2:57  
DRYING: 30:00  
TOTAL CYCLE: 01:00:34  
CYCLE COMPLETE

02/23/2021 12:20 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01