

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002085  
13:36 23/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:19  
PRESSURIZING 2:21  
STERILIZING 3:29  
136.5 C 225kPa 3:29  
136.2 C 224kPa 3:59  
136.3 C 225kPa 4:29  
136.3 C 224kPa 4:59  
136.4 C 225kPa 5:29  
136.2 C 223kPa 5:59  
136.3 C 224kPa 6:29  
136.4 C 226kPa 6:59  
136.4 C 226kPa 6:59  
VENTING 6:59  
AIR DRYING 7:51  
CYCLE COMPLETE 7:55  
-----

Digital Signature #  
C93FC1B0FE238836  
-----