
Midmark M11 - v1.0.5
Total Cycles: 2237

Name: RITTER2
Sterilizer ID: V2313228
04/12/2023 08:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.4 F 0.0
2:00 139.9 F 1.7
4:00 150.8 F 4.4
6:00 176.6 F 8.4
8:00 198.4 F 13.4
10:00 216.1 F 7.1
12:00 230.3 F 12.3
14:00 233.6 F 9.5
16:00 236.8 F 9.3
18:00 247.5 F 14.7
20:00 259.8 F 21.5
22:00 270.9 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.7 F 30.0
2:00 271.6 F 29.2
3:00 272.3 F 30.1
4:00 272.2 F 29.6

Min 271.3 F 29.2
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.0 F 0.0
5:00 233.3 F 0.0
10:00 230.2 F 0.0
15:00 225.2 F 0.0
20:00 224.0 F 0.0
25:00 219.7 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 22:46
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:58
CYCLE COMPLETE

04/12/2023 09:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____